

DOCUMENT ID:

**3310 Appendix T2
Operational Safety Procedure Form**

Click
for
OSP/TOSP Instructions

Serial Number: PHY-11-047-OSP / PHY-11-016-OSP

(Assigned by [ESH&Q Document Control](#) x7277)

☒ **OSP** ☐ **TOSP**

*Attach the Task Hazard Analysis (THA) related to this procedure

Click
For Word Doc

Issue Date: 12/12/2011

Expiration Date: 12/12/2014

(No more than three years from Issue Date except TOSP which is three months from issue date)

Title: Hall A Tungsten Calorimeter Operation

Location: HALL A

Risk classification

(See [ESH&Q Manual Chapter 3210 Appendix T3 Risk Code Assignment](#).)

Without mitigation measures (3 or 4): 3

With mitigation measures in place (0, 1, or 2): 1

Document Owner(s): Alexandre Camsonne

Date: 11/11/2011

Supplemental Technical Validations:

Hazard Reviewed (per [ESH&Q Manual 2410-T1](#)):

Subject Matter Experts Signature:

Date:

Radiation Hazards

Pavel Degtiarenko or Keith Welch 12/14/11

GENERAL EHS (BENT MANZLAK) Bent Manzlake 12-14-11

Approval Signatures:

Print

Signature

Date:

Division Safety Officer: Javier Gomez (Acting)

Department or Group Head: Robert Michaels

Safety Warden of Area: John Segal

Other Approval(s): Arne Freyberger

Document History:

Revision:

Reason for revision or update:

Serial number of
superseded document

1

First revision

1

Distribution: Copies to: affected area, authors, Division Safety Officer, ESH&Q Document Control

After expiration: Forward original and log sheet of trained personnel to ESH&Q Document Control.

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05/12	0	1 of 7

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

DOCUMENT ID:

3310 Appendix T2 Operational Safety Procedure Form

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

1. Purpose of the Procedure

The purpose of this document is to describe how operate the tungsten calorimeter in Hall A .

2. Scope – include operations, people, and/or areas where procedure applies

The Hall A beam calorimeter is a device to precisely calibrate the beam current monitors, so that the integrated charge may be known to better than 1% precision at low currents around 1 μ A. This is needed to measure precise cross sections even in this current range, for measurements such as in the low Q measurements of g2p and ep elastic scattering.

3. Description of the Facility: (include floor plans and layout of a typical experiment or operation)

The heart of the Hall A beam calorimeter is a tungsten-copper slug 16 cm in diameter and 16 cm long. The beam enters the calorimeter along its central, symmetry axis, and is stopped. The beam energy is almost entirely absorbed by the calorimeter around 0.4% of the energy escapes in the form of photons and neutrons. The measured temperature rise of the calorimeter determines the total energy absorbed. Using the beam energy, the total beam charge can be calculated, and the Hall A beam current monitors (BCMs) calibrated to agree. The calorimeter slug has three positions. First, it may be out of the beam on a cooling plate. Second, it may be out of the beam off the cooling plate; this position is used to monitor the slug temperature for an extended period, to determine the energy deposited in an exposure. Third, it may be in the beam. If the slug is moved when beam is on, it will trip the beam. The calorimeter was relocated in the Hall A beam line for the g2p experiment, from the center of the superharp girder to about a foot after the slow raster which is itself 2 feet apart from the usual Hall A fast raster. Controlling electronics are located on the floor directly below the calorimeter. In ordinary use, the equipment uses only low voltage, low power electronics.

4. Authority and Responsibility:

4.1 Who has authority to implement/terminate

John Segal / Ed Folts

4.2 Who is responsible for key tasks

Alexandre Camsonne

5. Who analyzes the special or unusual hazards (See [ES&H Manual Chapter 3210 Appendix T1 Work Planning](#),

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05/12	0	2 of 7

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

DOCUMENT ID:

3310 Appendix T2 Operational Safety Procedure Form

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

Control, and Authorization Procedure)

Pavel Degtiarenko

6. Personal and environmental hazard controls including:

6.1 Shielding

6.2 Interlocks

Interlocked with Fast Shut Down system

6.3 Other

7. Monitoring systems

Temperature with RTDs

8. Ventilation

9. List of safety equipment (i.e: personal protective equipment or special tools)

10. Associated administrative procedures

Hall A Calorimeter beam delivery procedure

11. Operating guidelines

The detailed normal operations of the calorimeter are precisely defined in Accelerator Division Operational Procedures and Safety Protocols, included as appendices to this document. A calorimeter expert or other authorized person will be present for all initial commissioning operations.

- Normal operations

The detailed normal operations of the calorimeter are precisely defined in Accelerator Division Operational Procedures and Safety Protocols, included as appendices to this document. A calorimeter expert or other authorized person will be present for all initial commissioning operations.

- Turning Off The Calorimeter

The calorimeter is turned "off" by MCC moving it to the cooling plate position or to the out of beam and off cooling plate position. In these positions, the calorimeter temperature is monitored, and the cooling water system continues to operate. Turning off the water cooler system requires access to the Hall, and

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05012	0	3 of 7

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

DOCUMENT ID:

3310 Appendix T2 Operational Safety Procedure Form

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

should only be done if the system will not be used for an extended time.

- Working on or Removing the Calorimeter

Since any exposure to beam makes the calorimeter radioactive, any work on the system subsequent to beam exposure must be coordinated with radcon. In particular, since the water in the cooler system might be activated, it should be treated as any potentially activated, contained water is treated.

12. Notification of Affected Personnel (How and Who)

Radiation Control Group if work on actual calorimeter

13. List of steps required to execute the procedure from start to finish.

1. Prerequisite

Current request must be less than 90% of Max Current I_{Limit} computed by the software.

2. Exposure time requested Δt_{max} must be less than maximum computed exposure time.

- Determine beam delivery configuration by establishing requested current using the Hall A BCM and load golden orbit established during commissioning.
- Secure beam to Hall A-Beam Sync/Close the Hall A slit.
- Mask appropriate FSD
- Enable Hall A Current lock
- Disable Ramp Feature of Hall A Current Lock
- Go to Empty target and set Fast and Slow Raster size to obtain 2mm x 2 mm surface at the calorimeter. Check with viewer.
- Calculate current and time exposure limits using the following equations:

$$I_{Limit} (\mu A) = 2500 \text{ (Watts)} / E_{beam} \text{ (MeV)}$$

$$\Delta t_{max} (s) = 150,000 \text{ J} / (I_{beam} (\mu A) \cdot E_{beam} \text{ (MeV)})$$

2. Setup Procedure

- Open Calorimeter Control Screen
- Select the "In Beam" button on the Control Page
- Verify that the Position Read back is "In Beam"
- Unmask FSD masked in step 5 of the Prerequisites
- Set/Verify Thermal Limit (50°C)
- Verify current limit has been entered into Max Juice and BLA system
- Adjust target ion chamber as necessary to perform measurement
- Set Timer with exposure time requested by hall.

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05/12	0	4 of 7

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

DOCUMENT ID:

3310 Appendix T2 Operational Safety Procedure Form

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

3. Exposure Procedure

The operator should manually discontinue beam delivery during this procedure

- If the current breaches the limit

- If any of the Resistive Temperature Devices (R. T. D.'s) reaches the thermal limit or fails. Note: The calorimeter will move to the cooling plate if piochal fails.

- When the countdown timer reaches zero

The current calibration software should send the limit to the BLA system and Max Juice

- Establish beam to the calorimeter

- Software automatically starts timer. Timer should be paused for FSD trips.

- Monitor current after establishing beam

4. Measurement Procedure

- Secure beam to Hall A-Beam Sync/Close the Hall A slit.

- Mask appropriate FSD

- Select the "Equilibrate" button on the control screen

- Verify position is "Equilibrate"

- Unmask FSD and restore target ion chamber.

- Physics to Hall A may continue at this time.

5. Back Out procedure

- Secure beam to Hall A-Beam, Sync/Close the Hall A slit.

- Mask appropriate FSD

- Select the "Cooling Plate" button on the control screen

- Verify position is "Cooling Plate"

- Unmask FSD

- Make an e-log indicating that Hall A cal-cal measurement is complete or aborted

14. Back out procedures, i.e., steps necessary to restore the equipment/area to a safe level.


Move back to out of beam position

15. Special environmental control requirements:

Work on detector or disposal requires Radcon supervision

16. Environmental Impacts (See [EMP-04 Project/Activity/Experiment Environmental Review](#))

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05012	0	5 of 7

		TITLE:	ES&H Manual
DOCUMENT ID:	3310 Appendix T2 Operational Safety Procedure Form		

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

17. Abatement Steps – Secondary Containment, or Special Packaging requirements
If detector is taken out it will have to be stored in a radiation area
18. Training requirements
Read and sign this OSP
19. Unusual/Emergency procedures e.g., Injury, Fire, Loss of power
20. Instrument calibration requirements, e.g., safety system/device recertification, RF probe calibration
21. Inspection schedules
22. References/Associated Documentation
23. List of Records Generated (Include Location / Review and Approved procedure)

Authorized/Trained Individuals

Print Name/Signature	Date
Alexandre Camsonne	
Arne Freyberger	
Mahmoud Ahmad	
Pengjia Zhu	

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05/12	0	6 of 7

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

DOCUMENT ID:

3310 Appendix T2
Operational Safety Procedure Form

Serial Number:

(Assigned by [ESH&Q Document Control](#) x 7277)

Print Name/Signature	Date

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05012	0	

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

ISSUING AUTHORITY	APPENDIX AUTHOR	APPROVAL DATE	EFFECTIVE DATE	EXPIRATION DATE	REV.	Page 7 of 7
ESH&Q Division	Harry Fanning	10/05/09	01/01/10	10/05012	0	

This document is controlled as an on line file. It may be printed but the print copy is not a controlled document. It is the user's responsibility to ensure that the document is the same revision as the current on line file. This copy was printed on 12/13/2011.

Calorimeter Measurement Procedure -- 1-1

Hall A Calorimeter Measurement Procedure

☐ Establish requested Current to Hall A

Setup for Exposure

- ☐ Secure beam to Hall A
- ☐ Set Calorimeter to "In Beam"
- ☐ Configure Current Lock and BLA
- ☐ Configure raster X
- ☐ Wait for prompt before sending CW min
- ☐ Wait for prompt to continue (min)

Move to Equilibrium Plate

- ☐ Secure Beam to Hall A
- ☐ Set Calorimeter to "Equilibrate"
- ☐ Restore Current Lock and BLA
- ☐ Restore Raster
- ☐ Wait for prompt to backout min

Backout

- ☐ Secure Beam to Hall A
- ☐ Set Calorimeter to "Cooling Plate"

	Max	Request	Readback
Current (uA)	1.45	<input type="text" value="1.305"/>	<input type="text" value="0.00"/>
Time (min)	1.00	<input type="text" value="0.57"/>	<input type="text" value="0.57"/>
Temp (C)	50	<input type="text" value="50"/>	

Slug RTD Temps (C)		Wall RTD Temps (C)	
Average	<input type="text" value="20.542"/>	Average	<input type="text" value="22.256"/>
P11042	<input type="text" value="20.275"/>	1	<input type="text" value="22.538"/>
P11044	<input type="text" value="20.800"/>	2	<input type="text" value="21.975"/>
P11047	<input type="text" value="20.425"/>		
P11048	<input type="text" value="20.725"/>		
P11051	<input type="text" value="20.350"/>		
P11055	<input type="text" value="20.675"/>		

EDM Screen

Help

Quit