

# Technical Report



Subject: JLab Q1 JSA-09-C1564                      Author: FD  
Instrumentation Design Summary  
Milestone D-13 - Issue 4

---

Ref No.: P0425SPM                                      Date: 4<sup>th</sup> June 2013

---

Dist: PB, SL (JLab)                                      CC: SRM, PNP

---

This document has been updated to Issue 4 with comment from JLab ref email dated 23 November 2010.

---

The Instrumentation design details provided in the JSA-09-C1564 contract documentation includes the following documents and descriptions for subcontractor supplied parts:

- Temperature sensors as defined in section 8.1 of the SHMS technical specification dated 14<sup>th</sup> July 2008 and wiring with pin outs defined on drawing 67125-E-00110\_RevF.
- Strain Gauges for cold to warm support as defined in section 8.6 of the SHMS technical specification dated 14<sup>th</sup> July 2008 and wiring with pin outs defined on drawing 67125-E-00112\_RevA.
- Voltage taps for magnet and current leads as defined in section 8.7 of the SHMS technical specification dated 14<sup>th</sup> July 2008 and wiring with pin outs defined on drawing 67125-E-00111\_RevA and overall schematic 67125-D-00113.

Each temperature and strain sensor has a duplicate redundant sensor to allow use of an alternative if the primary stops functioning.

Generally the sensor layout is understood by Scientific Magnetics and no changes would be proposed except those outlined in this document which require a response from JLab. There are several questions arising from review of the proposed instrumentation scheme before implementation directly which are summarised below as requested actions for JLab:

## Temperature Sensors

- JLab to confirm they will furnish fully pre-wired 41 pin connector with flying lead to the CCR contractor who will use the appropriate leads, leaving the remaining flying leads for use by Scientific Magnetics when the CCR is delivered. (see JLab email dated 23 November 2010)
- JLab to confirm that all sensor requirements for the magnet assembly are shown on tables below and that connection pins and responsibilities are correct. (see JLab email dated 23 November 2010)
- Exact location for all temperature sensors is to be clarified.

## Strain gauge sensors:

- The 8 strain gauges in full bridge configuration per rod are understood by SM and will be implemented as defined by JLab.

**Recommendations for voltage taps**

- JLab to confirm they will furnish fully pre-wired 41 pin connector with flying lead to the CCR contractor who will use the appropriate leads, leaving the remaining flying leads for use by Scientific Magnetix when the CCR is delivered. (see JLab email dated 23 November 2010)

**Temperature Sensors Ref 67125-E-00110**

**PT100 temperature sensors:**

We propose to use our standard PT100 sensor which conforms with IEC751 class B which is equivalent to the required DIN 43760 (Class B). A data sheet is attached in Appendix 1 for information. It has successfully been used on superconducting magnet systems for measurement of temperatures above 10K (with the use of a suitable calibration curve).

It is noted that these sensors have been configured in three wire mode, mainly due to the lack of pins on the CCR connectors.

**Carbon Sensors:**

We propose to use our standard carbon-ceramic sensor TVO type D2 which is 3 point calibrated, a calibration curve to interpolate between these 3 points will be provided. A data sheet on this sensor is attached in Appendix 2 for information:

The responsibility for PT100 and Carbon temperature sensor wiring is shown in the following four tables

Primary Sensors as defined in 67125-E-00110 RevF

Sensor ID	Location	Connector	Pins	Fit and Wiring Responsibility
CG-1 T_Coil_1	Magnet coil 1	1	1,2,3,4	Scientific Magnetics
CG-2 T_Coil_2	Magnet coil 2	1	5,6,7,8	Scientific Magnetics
CG_3 T_Coil_3	Magnet coil 3	1	9,10,11,12	Scientific Magnetics
CG_4 T_Coil_4	Magnet coil 4	1	13,14,15,16	Scientific Magnetics
CG_5 T_CL_N_C	Current lead negative in CCR	1	17,18,19,20	Scientific Magnetics
CG_6 T_CL_P_C	Current Lead positive in CCR	1	21,22,23,24	Scientific Magnetics
PT102-1 PT_Yoke_1	Yoke	1	25,26,27	Scientific Magnetics
PT102-2 PT_Yoke_2	Yoke	1	28,29,30	Scientific Magnetics
PT102-3 PT_Yoke_3	Yoke	1	31,32,33	Scientific Magnetics
PT102-4 PT_Yoke_4	Yoke	1	34,35,36	Scientific Magnetics
Short		1	40, 41	Scientific Magnetics

Redundant Sensors as defined in 67125-E-00110 Rev F

Sensor ID	Location	Connector	Pins	Fit and Wiring Responsibility
CG-1-R T_Coil_1	Magnet coil 1	2	1,2,3,4	Scientific Magnetics
CG-2-R T_Coil_2	Magnet coil 2	2	5,6,7,8	Scientific Magnetics
CG_3-R T_Coil_3	Magnet coil 3	2	9,10,11,12	Scientific Magnetics
CG_4-R T_Coil_4	Magnet coil 4	2	13,14,15,16	Scientific Magnetics
CG_5-R T_CL_N_C	Current lead negative in CCR	2	17,18,19,20	Scientific Magnetics
CG_6-R T_CL_P_C	Current Lead positive in CCR	2	21,22,23,24	Scientific Magnetics
PT102-1-R PT_Yoke_1	Yoke	2	25,26,27	Scientific Magnetics
PT102-2-R PT_Yoke_2	Yoke	2	28,29,30	Scientific Magnetics
PT102-3-R PT_Yoke_3	Yoke	2	31,32,33	Scientific Magnetics
PT102-4-R PT_Yoke_4	Yoke	2	34,35,36	Scientific Magnetics
Short		2	40,41	Scientific Magnetics

Primary Sensors as defined in 67125-E-00110 Rev. F

Sensor ID	Location	Connector	Pins	Fit and Wiring Responsibility
PT102_6 PT_N2_OUTER_TOP	Magnet outer nitrogen shield top	A	1,2,3	Scientific Magnetics
PT102_7 PT_N2_OUTER_BOTTOM	Magnet outer nitrogen shield bottom	A	4,5,6	Scientific Magnetics
PT102_8 PT_N2_BORE_TOP	Magnet bore nitrogen shield top	A	7,8,9	Scientific Magnetics
PT102_9 PT_N2_BORE_BOTTOM	Magnet bore nitrogen shield bottom	A	10,11,12	Scientific Magnetics
PT102_5 PT_N2_IN	Nitrogen feed line	B	1,2,3	CCR Contractor
PT102_10 PT_N2_RETURN	Nitrogen return line	B	4,5,6	CCR Contractor
DIODE-1 TD_HE_COOLDOWN	He pipework	B	7,8,9,10	CCR Contractor
DIODE-2 TD_HE_SUPPLY	He pipework	B	11,12,13,14	CCR Contractor
DIODE-3 TD_HE_COLD_RETURN	He pipework	B	15,16,17,18	CCR Contractor
DIODE-4 TD_HE_WARM_RETURN	He pipework	B	19,20,21,22	CCR Contractor
CG_7 T_HE_RESV	Helium reservoir	B	23,24,25,26	CCR Contractor
Short		B	40,41	CCR Contractor

#### Redundant Sensors as defined in 67125-E-00110

Sensor ID	Location	Connector	Pins	Fit and Wiring Responsibility
PT102_6_R PT_N2_OUTER_TOP	Magnet outer nitrogen shield top	A	13,14,15	Scientific Magnetics
PT102_7_R PT_N2_OUTER_BOTTOM	Magnet outer nitrogen shield bottom	A	16,17,18	Scientific Magnetics
PT102_8_R PT_N2_BORE_TOP	Magnet bore nitrogen shield top	A	19,20,21	Scientific Magnetics
PT102_9_R PT_N2_BORE_BOTTOM	Magnet bore nitrogen shield bottom	A	22,23,24	Scientific Magnetics
Short		A	40,41	Scientific Magnetics
PT102_5_R PT_N2_IN	Nitrogen feed line	C	1,2,3	CCR Contractor
PT102_10_R PT_N2_RETURN	Nitrogen return line	C	4,5,6	CCR Contractor
DIODE-1_R TD_HE_COOLDOWN	He pipework	C	7,8,9,10	CCR Contractor
DIODE-2_R TD_HE_SUPPLY	He pipework	C	11,12,13,14	CCR Contractor
DIODE-3_R TD_HE_COLD_RETURN	He pipework	C	15,16,17,18	CCR Contractor
DIODE-4_R TD_HE_WARM_RETURN	He pipework	C	19,20,21,22	CCR Contractor
CG-7_R T_HE_RESV	Helium reservoir	C	23,24,25,26	CCR Contractor
Short		C	40,41	CCR Contractor

#### Recommendations for Temperature Sensors

For connectors A, 1 and 2 there is a mix of responsibility for wiring the connector, and the pins to be used by the CCR contractor and Scientific Magnetics are not separated by pin such that it is highly likely that the current configuration will be damaged during the two step wiring operation because the pin selection is not aligned to the construction by two separate contractors.

In order to prevent this it is proposed that JLab consider the following:

---

Scientific Magnetics is a trading name of Space Cryomagnetics Ltd. Registered in England and Wales. Company No. 3950388.

VAT Registered No 749 9266 70

- JLab furnish a completely pre-wired plug to the CCR contractor with sufficient flying lead that both the CCR contractor and Scientific Magnetics can join their sensors to the wires, by use of a connector or soldered joint in the relevant vacuum space as appropriate. For avoidance of doubt this means the CCR contractor will use the JLab furnished pre wired 41 way connector to attach its sensors via appropriate pins, leaving the remaining flying leads accessible in the CCR base so that Scientific Magnetics can use the assigned leads for wiring its sensors later. This avoids any possibility of damaging leads already connected to the 41 way connector.

Note: Exact location for all temperature sensors is to be clarified.

#### **Strain Gauges Ref 67125-E-00012 Rev A**

##### **Strain gauge sensors:**

The wiring for the strain gauge sensors for the top vertical supports are detailed in the following table. The strain gauges are bonded to the support rods on a 20mm diameter section such that they make up a full bridge transducer configuration with four wires from each bridge connected to an 8 way lemo connector. The transducer is configured to measure tension along the axis of the rod (effectively  $2.6\epsilon$ ) and inherently corrects for bending of the rod. Each bridge has a redundant / spare installation and this is accessed on another four pins on the 8 way connector.

Strain Gauge ID	Location	Connector	Pins	Fit and Wiring Responsibility
SG1/1	Active strain gauge bridge - top rod	A	1-4	Scientific Magnetics
SG1/2	Redundant strain gauge bridge - top rod	A	5-8	Scientific Magnetics
SG2/1	Active strain gauge bridge - top rod	B	1-4	Scientific Magnetics
SG2/2	Redundant strain gauge bridge - top rod	B	5-8	Scientific Magnetics
SG3/1	Active strain gauge bridge - top rod	C	1-4	Scientific Magnetics
SG3/2	Redundant strain gauge bridge - top rod	C	5-8	Scientific Magnetics
SG4/1	Active strain gauge bridge - top rod	D	1-4	Scientific Magnetics
SG4/2	Redundant strain gauge bridge - top rod	D	5-8	Scientific Magnetics

The wiring for the strain gauge sensors for the side supports are detailed in the following table. The gauge configuration is identical to that on the vertical rods.

Strain Gauge ID	Location	Connector	Pins	Fit and Wiring Responsibility
SG5/1	Active strain gauge bridge - side rod	E	1-4	Scientific Magnetics
SG5/2	Redundant strain gauge bridge - side rod	E	5-8	Scientific Magnetics
SG6/1	Active strain gauge bridge - side rod	F	1-4	Scientific Magnetics
SG6/2	Redundant strain gauge bridge - side rod	F	5-8	Scientific Magnetics
SG7/1	Active strain gauge bridge - side rod	G	1-4	Scientific Magnetics
SG7/2	Redundant strain gauge bridge - side rod	G	5-8	Scientific Magnetics
SG8/1	Active strain gauge bridge - side rod	H	1-4	Scientific Magnetics
SG8/2	Redundant strain gauge bridge - side rod	H	5-8	Scientific Magnetics



### Voltage Taps Ref 67125-E-00111 and 67125-D-00113

Voltage taps and the responsibility for making them are shown in the following tables

Flying cable

Pot Tap ID	Location	Connector	Pins	Fit and Wiring Responsibility
POT TAP I+U	+ current lead top	N/A	N/A	Scientific Magnetics
POT TAP I-U	- current lead top	N/A	N/A	Scientific Magnetics
POT TAP I+U_R	Magnet outer nitrogen shield bottom	N/A	N/A	Scientific Magnetics
POT TAP I-U_R	Magnet bore nitrogen shield top	N/A	N/A	Scientific Magnetics

#### 41 pin connector - Primary pot taps

Pot Tap ID	Location	Connector	Pins	Fit and Wiring Responsibility
POT TAP I+M1	+ current lead lap joint in CCR top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	1	CCR Contractor
POT TAP I+M2	+ current lead lap joint in CCR bottom	Item 10 on dwg 67125-E-00111 Rev A sheet 2	2	CCR Contractor
POT TAP I+L1	+ current lead lap joint in interface top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	3	Scientific Magnetics
POT TAP I+L2	+ current lead lap joint in interface bottom / coil 1 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	4	Scientific Magnetics
POT TAP 1/1B	Coil 1 end/ coil 2 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	5	Scientific Magnetics
POT TAP 2/1A	Coil 1 end / Coil 2 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	6	Scientific Magnetics
POT TAP 2/1B	Coil 2 end / Coil 3 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	7	Scientific Magnetics
POT TAP 3/1A	Coil 2 end / Coil 3 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	8	Scientific Magnetics
POT TAP 3/1B	Coil 3 end / Coil 4 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	9	Scientific Magnetics
POT TAP 4/1A	Coil 3 end / Coil 4 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	10	Scientific Magnetics
POT TAP I-L2	- current lead lap joint in interface bottom / coil 4 end	Item 10 on dwg 67125-E-00111 Rev A sheet 2	11	Scientific Magnetics
POT TAP I-L1	- current lead lap joint in interface top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	12	Scientific Magnetics
POT TAP I-M2	- current lead lap joint in CCR top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	13	CCR Contractor
POT TAP I-M1	+/-current lead lap joint in CCR bottom	Item 10 on dwg 67125-E-00111 Rev A sheet 2	14	CCR Contractor

#### 41 pin connector - Redundant pot taps

Pot Tap ID	Location	Connector	Pins	Fit and Wiring Responsibility
POT TAP I+M1_R	+ current lead lap joint in CCR top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	15	CCR Contractor
POT TAP I+M2_R	+ current lead lap joint in CCR bottom	Item 10 on dwg 67125-E-00111 Rev A sheet 2	16	CCR Contractor
POT TAP I+L1_R	+ current lead lap joint in interface top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	17	Scientific Magnetics
POT TAP I+L2_R	+ current lead lap joint in interface bottom / coil 1 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	18	Scientific Magnetics
POT TAP 1/1B_R	Coil 1 end/ coil 2 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	19	Scientific Magnetics
POT TAP 2/1A_R	Coil 1 end / Coil 2 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	20	Scientific Magnetics
POT TAP 2/1B_R	Coil 2 end / Coil 3 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	21	Scientific Magnetics
POT TAP 3/1A_R	Coil 2 end / Coil 3 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	22	Scientific Magnetics
POT TAP 3/1B_R	Coil 3 end / Coil 4 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	23	Scientific Magnetics
POT TAP 4/1A_R	Coil 3 end / Coil 4 start	Item 10 on dwg 67125-E-00111 Rev A sheet 2	24	Scientific Magnetics
POT TAP I-L2_R	- current lead lap joint in interface bottom / coil 4 end	Item 10 on dwg 67125-E-00111 Rev A sheet 2	25	Scientific Magnetics
POT TAP I-L1_R	- current lead lap joint in interface top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	26	Scientific Magnetics
POT TAP I-M2_R	- current lead lap joint in CCR top	Item 10 on dwg 67125-E-00111 Rev A sheet 2	27	CCR Contractor
POT TAP I-M1_R	+/-current lead lap joint in CCR bottom	Item 10 on dwg 67125-E-00111 Rev A sheet 2	28	CCR Contractor
Short		Item 10 on dwg 67125-E-00111 Rev A sheet 2	40, 41	CCR Contractor

#### Recommendations for voltage taps

For this connector there is a mix of responsibility for wiring the connector, and the pins to be used by the CCR contractor and Scientific Magnetics are not separated by pin such that it is highly likely that the current configuration will be damaged during the two step wiring operation because the pin selection is not aligned to the construction by two separate contractors. In order to prevent this it is proposed that JLab consider the following:

---

Scientific Magnetics is a trading name of Space Cryomagnetics Ltd. Registered in England and Wales. Company No. 3950388.

VAT Registered No 749 9266 70

- JLab furnish a completely pre-wired plug to the CCR contractor with sufficient flying lead that both the CCR contractor and Scientific Magnetics can join their voltage taps to the wires, by use of a connector or soldered joint in the relevant helium space as appropriate.

## • Appendix 1 – PT100 Sensor data sheet

### Platinum Resistance Thermometry



Platinum Resistance Thermometers (PRT), constructed using thin film element

The sensor is platinum layered, laser trimmed and sits on a ceramic substrate, ideal for gas and surface measurements, giving rapid response due to low mass and good thermal transfer.

#### Pt100 Element, Thin Film Type (100 Ohm)

- Conforms to IEC751 Class B
- -50°C to +550°C temperature measurement range
- 10mm tails
- Ideal solution for surface mounting measurement applications
- Immersion applications where protected
- Fast response times
- Good vibration resistance
- Long-term stability

#### Specifications of element

- |                                       |                 |
|---------------------------------------|-----------------|
| • Temperature range                   | -50°C to +550°C |
| • Ice point resistance                | 100Ω            |
| • Fundamental interval (0°C to 100°C) | 38.5Ω nom.      |
| • Self heating                        | 0.005°C/mW      |
| • Thermal response                    | 0.1s            |
| • Stability                           | ±0.05%          |

#### Dimensions & Class

Dimensions L x W x H  
2 x 5 x 1.1mm  
Class B

#### RS Stock No

290-5070

RS Data 290-5070 12.11-08 GC

## Appendix 2 – TVO Carbon ceramic sensor data sheet



### MODEL TMI-A1 CCS

#### CCS Carbon-Ceramic Resistor

The Series TMI-A1 CCS, from the TVO family of resistors, is based on a carbon-ceramic composite construction which adheres to a single resistance versus temperature curve. They offer excellent performance and stability characteristics in magnetic field and high dose radiation environments. TMI-CCS resistors have little magneto-resistance in fields up to 6T over their range. No orientation dependence of resistor mounting relative to the magnetic field has been observed. TMI-CCS resistors are available in calibrated, sorted and un-calibrated forms featuring fast response times, excellent repeatability and high mechanical, thermal and radiation stability - all at reasonable cost.

#### CCS-A1 Specifications

**Useful Temperature Range**  
1.5K - 375K

**Maximum Operating Temperature**  
450K

**Nominal Resistivity (typical)**  
4363  $\Omega$  @ 4.2K  
1900  $\Omega$  @ 20K  
1238  $\Omega$  @ 77K  
1156  $\Omega$  @ 100K  
850  $\Omega$  @ 295K

#### Key Product Features

- excellent long-term stability ( $< 0.015\text{K}/15\text{yr}$ )
- thermal response  $< 1\text{ms}$  @ 4.2K
- high sensitivity 1000-1500  $\Omega/\text{K}$  @ 4.2K
- low magnetic field error  $< 1\%$  for  $B < 6\text{T}$
- low neutron-radiation error  $< 1\%$  for  $F < 10^{17}\text{n}/\text{m}^2$
- superior mechanical stability

#### Standard Configuration

Special glass coated carbon-ceramic matrix approximately 2mm x 8mm x 1mm (deep)

#### Available Models

Band	Temp. Range	$\Omega$ @ 4.2k	Sensitivity ( $\Omega/\text{K}$ )	Status
A1	1.5K - 375k	4000 - 5000	1000 - 1500	calibrated
A2	1.5K - 375k	3000 - 4000	600 - 1000	calibrated
B1	2.5K - 375k	3500 - 4000	800 - 1200	calibrated
B2	2.5K - 375k	2800 - 3500	500 - 800	calibrated
C1	4.2K - 375k	3500 - 4000	800 - 1200	calibrated
C2	4.2K - 375k	2800 - 3500	500 - 800	calibrated
D1	1.5K - 375k	4000 - 6000		selected
D2	4.2K - 375k	2600 - 4000		selected

#### Contact Information:

Tmi, Business Logistix Ltd,  
Fraternity House, 1 Harborne Road, Tackley, Oxford, UK, OX5 3BL  
Telephone: +44 (0) 1869 331108 Fax: +44 (0) 1869 331641  
web: <http://www.temati-uk.com/> email: [info@temati-uk.com](mailto:info@temati-uk.com)

Typical Sensitivity R/T ( $\Omega/\text{K}$ )											
$\Omega/\text{K}$	300	250	200	150	100	77	30	20	10	4.2	
$R_{4.2}=6300$	1	1.25	1.7	2.5	4	6	25	50	180	1135	
$R_{4.2}=5700$	0.85	1.1	1.5	2.4	3.8	6	20	40	165	1005	
$R_{4.2}=4500$	0.83	1	1.5	2.3	3.7	5	20	35	120	620	
$R_{4.2}=3500$	0.6	0.8	1	1.4	2.5	3.5	15	25	75	390	

