


PROFIBUS Pin-Outs


If your Alicat instrument was ordered with a PROFIBUS connection, please be sure to reference the following pinout diagram.

PROFIBUS



1.	Not Connected
2.	Power over Profibus Ground (optional)
3.	Rx/Tx Data-P
4.	Not Connected
5.	Data Ground
6.	Voltage Plus
7.	Power over Profibus 7 to 30 VDC (Optional)
8.	Rx/Tx Data-N
9.	Not Connected

Power/Serial



1.	RS-232 TX (output)
2.	Power In 7 to 30 VDC
3.	Ground (common)
4.	RS-232 RX (input)