





pH I²C



/**THIS CODE WILL WORK ON ANY ARDUINO**
 //This code has intentionally has been written to be overly lengthy and includes unnecessary steps.
 //Many parts of this code can be truncated. This code was written to be easy to understand.
 //Code efficiency was not considered. Modify this code as you see fit.
 //This code will output data to the Arduino serial monitor. Type commands into the Arduino serial monitor to control the EZO pH Circuit in I²C mode.

```

#include <Wire.h>
#define address 99

//enable I2C.
//default I2C ID number for EZO pH Circuit.

char computerdata[20];
byte received_from_computer=0;
byte serial_event=0;
byte code=0;
char ph_data[20];
byte in_char=0;
byte i=0;
int time=1400;
float ph_float;

//we make a 20 byte character array to hold incoming data from a pc/mac/other.
//we need to know how many characters have been received.
//a flag to signal when data has been recived from the pc/mac/other.
//used to hold the I2C response code.
//we make a 48 byte character array to hold incoming data from the pH circuit.
//used as a 1 byte buffer to store in bound bytes from the pH Circuit.
//counter used for ph_data array.
//used to change the delay needed depending on the command sent to the EZO Class pH Circuit.
//float var used to hold the float value of the pH.

void setup()
{
  Serial.begin(9600);
  Wire.begin();
}

//hardware initialization.
//enable serial port.
//enable I2C port.

void serialEvent(){
  received_from_computer=Serial.readBytesUntil(13,computerdata,20);
  computerdata[received_from_computer]=0;
  serial_event=1;
}

//this interrupt will trigger when the data coming from
//the serial monitor(pc/mac/other) is received.
//we read the data sent from the serial monitor
//(pc/mac/other) until we see a <CR>. We also count
//how many characters have been received.
//stop the buffer from transmitting leftovers or garbage.

void loop(){
  //the main loop.

  if(serial_event){
    if(computerdata[0]=='c'||computerdata[0]=='r')time=1400;
    else time=300;

    //if the serial_event=1.
    //if a command has been sent to calibrate or take a reading we
    //wait 1400ms so that the circuit has time to take the reading.
    //if any other command has been sent we wait only 300ms.

    Wire.beginTransmission(address);
    Wire.write(computerdata);
    Wire.endTransmission();

    //call the circuit by its ID number.
    //transmit the command that was sent through the serial port.
    //end the I2C data transmission.

    delay(time);

    //wait the correct amount of time for the circuit to complete its instruction.

    Wire.requestFrom(address,20,1);
    code=Wire.read();

    //call the circuit and request 20 bytes (this is more then we need).
    //the first byte is the response code, we read this separately.

    switch (code){
      case 1:
        Serial.println("Success");
        break;

        //switch case based on what the response code is.
        //decimal 1.
        //means the command was successful.
        //exits the switch case.

      case 2:
        Serial.println("Failed");
        break;

        //decimal 2.
        //means the command has failed.
        //exits the switch case.

      case 254:
        Serial.println("Pending");
        break;

        //decimal 254
        //means the command has not yet been finished calculating.
        //exits the switch case.

      case 255:
        Serial.println("No Data");
        break;

        //decimal 255.
        //means there is no further data to send.
        //exits the switch case.

    }

    while(Wire.available()){
      in_char = Wire.read();
      ph_data[i]= in_char;
      i+=1;
      if(in_char==0){
        Wire.endTransmission();
        break;
      }
    }

    //are there bytes to receive.
    //receive a byte.
    //load this byte into our array.
    //incur the counter for the array element.
    //if we see that we have been sent a null command.
    //reset the counter i to 0.
    //end the I2C data transmission.
    //exit the while loop.

    Serial.println(ph_data);
    serial_event=0;
    //print the data.
    //reset the serial event flag.
  }
}

//Uncomment this section if you want to take the pH value and convert it into floating point number.
/*
  ph_float=atof(ph_data);
*/
  
```

[Click here to download the *.ino file](#)