


FTB-4000 and FTB-5000 Series **Turbine Meters for Water**

Operator's Manual:



M0504/0201



General Description

Specifically designed for water billing applications, OMEGA: FTB-4000 and FTB-5000 Series Flowmeters are highly accurate and feature tamper-resistant, non-resettable totalizers. The large faces are easy to read and are fully rotatable for simplified mounting. FTB-4605 and FTB-4607 have no local indication of the flow rate or total.

Optional reed relay scaled pulse outputs allow for remote totalization (6 feet of cable included). For the FTB-4000 Series, the pulse output is factory installed and must be requested at the time of order.

The FTB-5000 Series units feature field installation of the reed relay or optical pickup pulse outputs. The optical pickup is used for use with flow rate frequency meters to provide remote rate indication (user DC power input required). FTB-5000 units are supplied standard with one reed relay for remote totalization; a second relay is optional.

All FTB flowmeters feature built-in strainers and are shipped complete with locking nuts, gaskets and coupling pieces. The units have a built-in flow finder which registers even a small trickle of water passing through the meter.

WARNING

THESE WATER FLOWMETERS HAVE PLASTIC INTERNAL PARTS THAT ARE RAPIDLY ATTACKED BY HYDROCARBON FLUIDS, SUCH AS GASOLINE, DIESEL FUEL, KEROSENE, AND SIMILAR MATERIALS. EVEN TRACE AMOUNTS OF THESE MATERIALS IN WATER WILL TEND TO ACCUMULATE IN THE PLASTIC PARTS LEADING TO COMPLETE FAILURE OF THE FLOWMETER. NOTE: WATER AND ETHYLENE GLYCOL MIXTURES TYPICALLY PRESENT NO PROBLEM WHEN USED WITH THESE WATER FLOWMETERS.

UNPACKING

Remove the packing list and verify that all equipment has been received. If there are any questions about the shipment, please call the OMEGA Customer Service Department.

Upon receipt of shipment, inspect the container and equipment for any signs of damage. Take particular note of any evidence of rough handling in transit. Immediately report any damage to the shipping agent.


The carrier will not honor any claims unless all shipping material is saved for their examination. After examining and removing contents, save packing material and carton in the event reshipment is necessary.

MOUNTING

The FTB-4000 Series (on y all 1/2" and 3/4") can be mounted either horizontally or vertically.

All FTB-4000 Series 1", 11/4", 11/2" and 2" horizontal only. FTB-5000X - FTB-5020 horizontal only.

FTB-5020X to FTB-5080 all can be mounted either horizontally or vertically







		FlowRate		Connections	Length	Height	Width	Pressure	Weight	Max
Model No.	Min.	GPM Cont.	Max	MNPT	(in.)	(in.)	(in.)	Loss (psi)	<u>(1)</u>	Temp
FTB-4005	0.15	9.9	13.2	1/2"C	4.33	2.50	2.75		1 10	122 F
FTB-4007	0.20	11.0	20.0	3/4/C	5.10	2.50	2.75		140	122°F
FTB-4105	0.15	9.9	13.2	1/2"C	4.33	2.50	2.75		1.10	200'F
FTB-4105P+	0.15		13.2	1/2"C	4.33	*	2.75		1.10	200'F
FTB-4107	0.20	11.0	20.0	3/4"C	5.10	2.50	2.75		1.40	200'F
FTB-4107P+	0.20	11.0	20.0	3/4"C	5.10	*	2.75		1.40	200'F
FTB-4110	0.50	26.4	52.8	1"C	10.25	4.75	4.00	2.00	5.50	200'F
FTB-4110P+	0.50	26.4	52.8	1"C	10.25	4.75	4.00	2.00	5.50	200'F
FTB-4112	0.50	26.4	52.8	1.25°C	10.25	4.75	4.00	2.00	5.50	200 F
FTB-4112P+	0.50	26.4	52.8	1.25°C	10.25	4.75	4.00	2.00	5.50	200'F
FTB-4115	0.80	44.0	88.0	1.5C	11.80	5.50	5.25	2.00	12.00	200'F
FTB-4115P+	0.80	44.0	88.0	1.5C	11.80	5.50	5.25	2.00	12.00	200'F
FTB-4605	0.15	9.9	13.0	1/2"C	4.33	2.00	2.75	3.60	1.20	190°F
FTB-4607	0.22	11.0	20.0	3/4"C	5.10	2.00	2.75	3.60	1 30	190,E
FTB-5000	0.13	9.9	13.2	1/2"C	4.33	4.50	3.25	2.90	2.20	248'F
FTB-5005	0.22	11.0	22.0	3/4"C	5.10	4.75	3.25	2.90	2.40	248'F
FTB-5010X	0.40	26.4	52.8	1"C	10.25	7.25	4.00	2.90	6.40	248'F
FTB-5015	0.70	43.9	87.2	1.5°C	11.80	8.10	5.25	3.60	11.25	248'F
FTB-5020	0.88	65.8	131.6	2FL	10.63	8.95	F	2.90	27.50	248'F
FTB-5020X	2.63	65.8	307.0	2FL	7.90	10.90	FL	0.15	31.50	248'F
FTB-5030	3.51	175.4	658.0	3FL	8.90	11.50	F	0.30	40.10	248'F
FTB-5040	5.26	263.2	790.0	4FL	9.85	12.20	F	0.40	43.70	248'F
FTB-5060	26.32	622.9	1535.0	149	11.80	14.10	4	0.30	71.60	248'F
FTB-5080	43.86	1096.5	2631.0	8FL	13.75	15.30	F	0.20	99.20	248'F
		P+ = Puise Output M	out Models							
		C = Union Coupling		connections with NPT threads	ads					
		FL = Flanged int	erconnection	FL = Flanged interconnection (conforms to ANSI B 16.5-150 lbs.	SI B 16.5-150 IL)S.)				
	-	• Height increase	es by approx	* Height increases by approximately 2 inches with pulse output	vith pulse outpu					
		Optional reed re	lay for remot	Optional reed relay for remote totalization-order model FTB-RR	r model FTB-R	α				

SPECIFICATIONS (CONT'D)

ACCURACY:

4600 Series: From 10% of cont. to max. flow: $\pm 1.5\%$ of reading

Below 10% of cont. flow: \pm 2% of reading

4000, 5000 Series: From 20% of cont. to max. flow: \pm 1% of reading

Below 20% of cont. flow: ± 3% of reading

MAX. TEMPERATURE:

 4000 Series:
 122°F (50°C)

 4100 Series:
 190°F (87.7°C)

 4600 Series:
 190°F (87.7°C)

5000 Series: 248°F (120°C)

MAX. PRESSURE:

4000, 4600 Series: 150 PSI **5000 Series:** 250 PSI

PULSE OUTPUTS: Reed relay

4100P Series: 1 gal./pulse

 5010 - 5020:
 1 or 10 gal./pulse

 5020X - 5040:
 10 gal./pulse

 5060 &: 5080:
 100 gal./pulse

FTB4600 HIGH RESOLUTION PULSE OUTPUT:

FTB4605: 151.4 pulses/gal.

FTB4607: 75.7 pulses/gal. Requires 6-16 VDC @ 10mA max power;

output requires pull-up to positive DC voltage.

(OPTICAL PICKUP (= MIN. INCREMENT):

5010 -5020X:0.5 gal./pulse5030 &: 5040:1 gal./pulse5060 &: 5080:5 gal./pulse

MAX. READING (GAL.):

 4000 Series:
 100,000,000

 5010 - 5040:
 10,000,000

 5060 - 5080:
 100,000,000

HOUSING:

 4000, 4100, 5005 - 5015:
 Cast brass

 5020 - 5080:
 Cast iron

 4600 Series:
 Brass

BEARING: Ceramic/sapphire

TURBINE: High temperature thermoplastic/fiberglass (polyimide)


FLOW TRANSFER: Ceramic magnet

OTHER WETTED PARTS: Stainless steel, polypropylene, EPDM O-ring

WIRING DIAGRAM (FTB4605/FTB4607 ONLY)

NOTEThe wiring scheme illustrated below may be of two versions Follow the appropriate wiring scheme.

LEAD	KEY	COLOR SCHEME
6-16VDC	Α	BROWN
PULSE SIGNAL HI	В	GREEN
POWER GROUND PULSE LO	С	WHITE





OMEGAnet® Online Service omega.com

FAX: (203) 359-7700

Internet e-mail info@omega.com

Germany/

Austria:

United

Servicing North America:

One Omega Drive, Box 4047 Stamford, CT 06907-0047 Tel: (203) 359-1660

ISO 9001 Certified

e-mail: info@omega.con

Canada:

976 Bergar Laval (Quebec) H7L 5A1, Canada Laval (Quebec) H7L 5A1, Canada FAX: (514) 856-6886

e-mail: info®omega ca

For immediate technical or application assistance:

U.S.A. and Canada:

U.S.A.:

Sales Service: 1-800-826-6342/1-800-TC-OMEGA^e Customer Service: 1-800-622-2378/1-800-622-BEST^e Engineering Service: 1-800-872-9436/1-800-USA-WHEN*

En Español: (001) 203-359-7803 e-mail:espanol@omega.com Mexico:

FAX: (001) 203-359-7807

info@omega.com.mx

Servicing Europe:

Frystatska 184, 733 01 Karviná, Czech Republic TEL: +420 (0)59 6311899 Czech Republic:

FAX: +420 (0)59 6311114 Tol Free: 0800-1-66342

e-mail: info@omegashop.ca Daimlerstrasse 26, D-75392 Deckenpfronn, Germany

Tel: +49 (0)7056 9398-0

FAX: +49 (0)7056 9398-29

Toll Free in Germany: 0800 639 7678

e-mail: info@omega.de

One Omega Drive, River Bend Technology Centre

Kingdom: Northbank, Irlam, Manchester ISO 9002 Certified M44 5BD United Kingdom

Tel: +44 (0)161 777 6611 FAX: +44 (0)161 777 6622

Toll Free in United Kingdom: 0800-488-488

e-mail: sales@omega.co.uk

It is the policy of OMEGA Engineering, Inc. to comply with all worldwide safety and EMC/EMI regulations that apply. OMEGA is constantly pursuing certification of its products to the European New Approach Directives. OMEGA will add the CE mark to every appropriate device upon certification

The information contained in this document is believed to be correct, but OMEGA accepts no liability for any errors it contains, and reserves the right to alter specifications without notice.

WARNING: These products are not designed for use in, and should not be used for, human applications.

WARRANTY/DISCLAIMER

OMEGA ENGINEERING, INC. warrants this unit to be free of defects in materials and workmanship for a period of 13 months from date of purchase. OMEGA's WARRANTY adds an additional one (1) month grace period to the normal one (1) year product warranty to cover handling and shipping time. This ensures that OMEGA's customers receive maximum coverage on each

If the unit malfunctions, it must be returned to the factory for evaluation. OMEGA's Customer Service Department will issue an Authorized Return (AR) number immediately upon phone or written request. Upon examination by OMEGA, if the unit is found to be defective, it will be repaired or replaced at no charge. OMEGA's WARRANTY does not apply to defects resulting from any action of the purchaser, including but not limited to mishandling, improper interfacing, operation outside of design limits, improper repair, or unauthorized modification. This WARRANTY is VOID if the unit shows evidence of having been tampered with or shows evidence of having been damaged as a result of excessive corrosion; or current, heat, moisture or vibration; improper specification; misapplication; misuse or other operating conditions outside of OMEGA's control. Components in which wear is not warranted, include but are not limited to contact points, fuses, and triacs.

OMEGA is pleased to offer suggestions on the use of its various products. However, OMEGA neither assumes responsibility for any omissions or errors nor assumes liability for any damages that result from the use of its products in accordance with information provided by OMEGA, either verbal or written. OMEGA warrants only that the parts manufactured by the company will be as specified and free of defects. OMEGA MAKES NO OTHER WARRANTIES OR REPRESENTATIONS OF ANY KIND WHATS:OFVER, EXPRESSED OR IMPLIED, EXCEPT THAT OF TITLE, AND ALL IMPLIED WARRANTIES INCLUDING ANY WARRANTY OF MERCHANTABILITY AND FITNESS FOR A PARTICULAR PURPOSE ARE HEREBY DISCLAIMED. LIMITATION OF LIABILITY: The remedies of purchaser set forth herein are exclusive, and the total liability of OMEGA with respect to this order, whether based on contract, warranty, negligence, indemnification, strict liability or otherwise, shall not exceed the purchase price of the component upon which liability is based. In no event shall OMEGA be liable for consequential, incidental or special damages.

CONDITIONS: Equipment sold by OMEGA is not intended to be used, nor shall it be used: (1) as a "Basic Component" under 10 CFR 21 (NRC), used in or with any nuclear installation or activity; or (2) in medical applications or used on humans. Should any Product(s) be used in or with any nuclear installation or activity, medical application, used on humans, or misused in any way, OMEGA assumes no responsibility as set forth in our basic WARRANTY/DISCLAIMER language, and, additionally, purchaser will indemnify OMEGA and hold OMEGA narmless from any liability or damage whatsoever arising out of the use of the Product(s) in such a manner.

RETURN REQUESTS / INQUIRIES

Direct all warranty and repair requests/inquiries to the OMEGA Customer Service Department. BEFORE RETURNING ANY PRODUCT(S) TO OMEGA, PURCHASER MUST OBTAIN AN AUTHORIZED RETURN (AR) NUMBER FROM OMEGA'S CUSTOMER SERVICE DEPARTMENT (IN ORDER TO AVOID PROCESSING DELAYS). The assigned AR number should then be marked on the outside of the return package and on any correspondence.

The purchase is responsible for shipping charges, freight, insurance and proper packaging to prevent breakage in transit.

FOR WARRANTY RETURNS, please have the following information available BEFORE contacting OMEGA:

- 1. Purchase Crder number under which the product was **PURCHASE D**
- 2. Model and serial number of the product under warranty, and
- 3. Repair instructions and/or specific problems relative to the product.
- FOR NON-WARRANTY REPAIRS, consult OMEGA for current repair charges. Have the following information available BEFORE contacting OMEGA:
- 1. Purchase Order number to cover the COST of the repair,
- 2. Model and serial number of the product, and
- 3. Repair instructions and/or specific problems relative to

OMEGA's policy is to make running changes, not model changes, whenever an improvement is possible. This affords our customers the latest in technology and engineering.

OMEGA is a registered trademark of OMEGA ENGINEERING, INC.

Omeda is a registrest acceptance of the control of