REVISION HISTORY				
ZONE	REV	DESCRIPTION	DATE	APPROVED



TOP VIEW 1-6G



NOTES:
1. ALL WELDS TO BE VACUUM TIGHT TO 1 X 10^-6 atm cc /sec.
2. WELD JOINTS TO BE COLD-SHOCKED 3 TIMES.

SECT VIEW 1-6A



DETAIL VIEW 1-2E
SCALE FULL



DETAIL VIEW 1-2C
SCALE FULL

QTY. REQD	ITEM NO.	PART OR IDENTIFYING NO.	NOMENCLATURE OR DESCRIPTION	MATERIAL SPECIFICATION	NOTES
1	1	A00000-09-04	U TUBE JUNCTION	316 STAINLESS STEEL	-

PARTS LIST

EACH SHEET OF MULTI-SHEET DRAWINGS SHALL ALWAYS HAVE THE SAME REVISION LEVEL.

DOCUMENT CONTROL STAMP	DIM & TOL PER ASME Y14.5 UNLESS OTHERWISE SPECIFIED DIMENSIONS ARE IN INCHES TOLERANCES FRACTIONS ± N/A DECIMAL X ±.1 XX ±.01 XXX ±.005 ANGLES ±.50°	TRACKING NO. N/A	United States Department of Energy — Jefferson Lab, Thomas Jefferson National Accelerator Facility — Newport News Virginia
MATERIAL SEE PARTS LIST	THIRD ANGLE PROJECTION	APPROVALS / DATE: DRAWN J. MILLER 28JUL09; CHECKED 3-AUG-09; APPROVED 3-AUG-09; APPROVED 3-AUG-09	HALL A - GENL EQUIPMENT BEAMLINE - SC MOELLER CRYOGENIC DISTRIBUTION SYSTEM H-TUBE FOR MOELLER
FINISH MACHINED SURFACES 125 UNLESS OTHERWISE NOTED; DEBURR & BREAK ALL SHARP EDGES			SIZE E; DWG. NO. A00000-09-04-0100; REV. -
DO NOT SCALE DRAWING			SCALE 1:4; USED ON NEXT NO. NA; SHEET 1 OF 1